

**April 5, 2017**

**EPIPEN RECALL:**

**Drug Company Mylan is recalling the items/lots below** due to the result of two previously disclosed reports outside of the U.S. of failure to activate the device due to a potential defect in a supplier component. This recall is to the Consumer level. Affected product started shipping December 2015.

**Products:**

**EPIPEN 0.3MG 0.3ML 1:1M 2PK**

NDC: 49502050002

UPC: 34950250002

**Lot Numbers/Expiration Dates:**

6GM199 10/31/17;

6GM198 10/31/17;

6GM091 10/31/17;

6GM088 10/31/17;

6GM087 10/31/17;

6GM082 09/30/17;

6GM081 09/30/17;

6GM072 09/30/17;

5GM640 05/31/17;

5GM631 04/30/17;

**EPIPEN JR .15MG .3ML 1:2M 2PK**

NDC: 49502050102

UPC: 34950250102

**Lot Numbers/Expiration Dates:**

6GN215 09/30/17

5GN773 04/30/17

5GN767 04/30/17

Please immediately check your inventory, quarantine, discontinue distribution of the affected product, and contact your wholesaler for directions.

For more information visit: <http://www.mylan.com/en/epipenrecall>